



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/595,305

06/14/2006

Moshe Shoham

20763/0204292-US0

7119

7278 7590 01/04/2010

DARBY & DARBY P.C.

P.O. BOX 770

Church Street Station

New York, NY 10008-0770

EXAMINER

COMSTOCK, DAVID C

ART UNIT

PAPER NUMBER

3733

MAIL DATE

DELIVERY MODE

01/04/2010

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/595,305	Applicant(s) SHOHAM ET AL.	
	Examiner DAVID COMSTOCK	Art Unit 3733	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-25 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-25 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 06 April 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____. |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>19 September 2006</u> . | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-14 and 16-22 are rejected under 35 U.S.C. 102(b) as being anticipated by Hein et al. (5,674,221; cited by Applicant).

Hein et al. disclose the claimed invention including one or more bridges for use in surgery comprising at least a first support member 104 (along a first side, see, e.g., Fig. 14), a second support member (also labeled 104, but the separate member along the other side in Fig. 14), and a cross member (also labeled 104, but the separate members closest to the bone; see Fig. 14 and cf. Fig. 13) attached to the first and second support members at positions remote from ends of the support members (see, e.g., Figs. 12-15 and col. 2, line 33 - col. 3, line 30). The bridge is capable of moving with movement of the spine (e.g. either fixed and together with the spine or as by articulation during adjustment). The support members are attached to bone via either K-wires (pins) or clamps (see, e.g., col. 4, lines 54-58). With regard to statements of intended use with a spine, skull, pelvic bone, etc., and other functional statements, these do not impose any structural limitations on the claims distinguishable over Hein et al., which is a surgical orthopedic device capable of being used as claimed. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law

Art Unit: 3733

of anticipation does not require that the reference “teach” what the subject patent teaches, but rather it is only necessary that the claims under attack “read on” something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 15 and 23-25 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hein et al. (5,674,221; cited by Applicant) in view of Mittelstadt et al. (6,322,567; cited by Applicant).

Hein et al. discloses the claimed invention including support structure capable of accommodating (e.g., providing structural mounting support) for a surgical robot, but does not explicitly recite using a surgical robot. However, Mittelstadt et al. teach that a surgical system can be provided with a robot 10 in order to detect and track bone motion and improve the accuracy of a surgical procedure (see, e.g., Figs. 1-3 and column 1, lines (see, e.g. col. 2, lines 2-6). It would have been obvious to one having ordinary skill in the art at the time the invention was made to provide the system of Hein

Art Unit: 3733

et al. with a surgical robot, in view of Mittelstadt et al., in order to track motion of the bone and improve the accuracy of the surgical procedure.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to David Comstock whose telephone number is (571) 272-4710 (a detailed message should be left if Examiner is unavailable). If attempts to reach the Examiner by telephone or voicemail are unsuccessful, the examiner's supervisor, Eduardo Robert, can be reached at (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/David Comstock/
Examiner, Art Unit 3733

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733